

VI ByPass Syringe™

Red Device with back-end needle
VI-102-BC

Intended Use:

VI ByPass Syringe™ BC is compatible with direct vacuum collection via direct venipuncture or vascular access.

Directions for Use

STEP 1: Prep supplies- Following **Standard Aseptic Non-Touch Technique** attach the VI ByPass Syringe™ (Red) to a Butterfly extension or IV extension set with or without a needleless connector. An extension set must be used when placing a new or existing peripheral IV catheter.

STEP 2: Cannulate vessel, aspirate and sequester up to 4mls per policy for Blood or Blood Culture collection. Insert tubes/bottles over the back-end needle and collect sample.

STEP 4: Remove Vacuum tubes/bottles from back-end needle. Sequestered volume may be reinfused according to procedural policy.

STEP 5: Disconnect VI ByPass Syringe™ and discard per hospital policy.

| | | | | |
|---|--|---|---|------------------------------------|
|  REF | - CATALOG NUMBER |  STERILE |  R | - STERILIZED USING IRRADIATION |
|  LOT | - LOT NUMBER |  | | - DO NOT USE IF PACKAGE IS DAMAGED |
|  | - NON-PYROGENIC |  | | - DO NOT RESTERILIZE |
|  | - CONSULT OPERATING INSTRUCTIONS |  | Rx only | - RX ONLY |
|  | - DO NOT REUSE |  | USE BY: xx-xx-xxxx | |
|  | - LATEX FREE |  | | - LATEX FREE |
|  | Vascular Integrity 16661 Jamboree Rd. Irvine, CA 92606 |  | ICU Medical 951 Calle Amanecer San Clemente, CA 92673 | |

VI ByPass Syringe™

Red Device with back-end needle
VI-102-BC

Intended Use:

VI ByPass Syringe™ BC is compatible with direct vacuum collection via direct venipuncture or vascular access.

Directions for Use

STEP 1: Prep supplies- Following **Standard Aseptic Non-Touch Technique** attach the VI ByPass Syringe™ (Red) to a Butterfly extension or IV extension set with or without a needleless connector. An extension set must be used when placing a new or existing peripheral IV catheter.

STEP 2: Cannulate vessel, aspirate and sequester up to 4mls per policy for Blood or Blood Culture collection. Insert tubes/bottles over the back-end needle and collect sample.

STEP 4: Remove Vacuum tubes/bottles from back-end needle. Sequestered volume may be reinfused according to procedural policy.

STEP 5: Disconnect VI ByPass Syringe™ and discard per hospital policy.

| | | | | |
|---|--|---|---|------------------------------------|
|  REF | - CATALOG NUMBER |  STERILE |  R | - STERILIZED USING IRRADIATION |
|  LOT | - LOT NUMBER |  | | - DO NOT USE IF PACKAGE IS DAMAGED |
|  | - NON-PYROGENIC |  | | - DO NOT RESTERILIZE |
|  | - CONSULT OPERATING INSTRUCTIONS |  | Rx only | - RX ONLY |
|  | - DO NOT REUSE |  | USE BY: xx-xx-xxxx | |
|  | - LATEX FREE |  | | - LATEX FREE |
|  | Vascular Integrity 16661 Jamboree Rd. Irvine, CA 92606 |  | ICU Medical 951 Calle Amanecer San Clemente, CA 92673 | |